



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
09/997,449	11/30/2001	Shamim M. Malik	50623.134	3441

7590 09/26/2006

Squire, Sanders & Dempsey L.L.P.  
Suite 300  
One Maritime Plaza  
San Francisco, CA 94111

EXAMINER
----------

HO, UYEN T

ART UNIT	PAPER NUMBER
----------	--------------

3731

DATE MAILED: 09/26/2006

Please find below and/or attached an Office communication concerning this application or proceeding.

## Office Action Summary

Application No.

09/997,449

Applicant(s)

MALIK ET AL.

Examiner

(Jackie) Tan-Uyen T. Ho

Art Unit

3731

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

### Status

- 1) ☒ Responsive to communication(s) filed on 26 June 2006.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

### Disposition of Claims

- 4) ☒ Claim(s) 1-31 is/are pending in the application.
- 4a) Of the above claim(s) 17-30 is/are withdrawn from consideration.
- 5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.
- 6) ☒ Claim(s) 1-16 and 31 is/are rejected.
- 7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.
- 8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

### Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

### Priority under 35 U.S.C. § 119

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
- a) ☐ All b) ☐ Some \* c) ☐ None of:
1. ☐ Certified copies of the priority documents have been received.
  2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.
  3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

### Attachment(s)

- 1) ☒ Notice of References Cited (PTO-892)
- 2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)
- 3) ☐ Information Disclosure Statement(s) (PTO/SB/08)  
Paper No(s)/Mail Date \_\_\_\_\_
- 4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date. \_\_\_\_\_
- 5) ☐ Notice of Informal Patent Application
- 6) ☐ Other: \_\_\_\_\_

## **DETAILED ACTION**

### ***Response to Arguments***

1. Applicant's arguments, see remarks, filed 6/26/06, have been fully considered and are persuasive. Therefore, the rejection has been withdrawn. However, upon further consideration, a new ground(s) of rejection is made in view of Kamath et al. (6,335,029).

### ***Claim Rejections - 35 USC § 103***

2. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.


3. Claims 1-16 and 31 are rejected under 35 U.S.C. 103(a) as being unpatentable over Kamath et al. (6,335,029) in view of Davidson (5,415,704). Kamath et al. disclose all the limitations of the claims except for a presence of the material as claimed for constructing a stent body which including a metallic substance and a carbon deposit present at a molecular level within the metallic substance and at a depth within a surface of the stent. The material as claimed are well known in the art for use in making vascular implant. Davidson disclose the material as claimed, which including a metallic substance and a carbon deposit present at a molecular level within the metallic substance and at a depth within a surface of an implant in order to strengthen and harden an alloy surface to gain the advantage of improved abrasion resistance for long

Art Unit: 3731

term placement of an implant in a body. Therefore, it would have obvious to one having ordinary skill in the art at the time the invention was made to modify or substitute the base metallic material of Kamath et al. with the metallic material as disclosed by Davidson in order to gain the advantage of improved abrasion resistance for long term placement of the stent in a body.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to (Jackie) Tan-Uyen T. Ho whose telephone number is 571-272-4696. The examiner can normally be reached on MULTIFLEX Mon. to Sat..

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, ANHTUAN NGUYEN can be reached on 571-272-4963. The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.



(Jackie) Tan-Uyen T. Ho  
Primary Examiner  
Art Unit 3731

September 18, 2006